



## Study of UV-Visible Spectroscopy

*Abhishek Borkar<sup>1</sup>, Shubhada Bhopale<sup>2</sup>, RaviKumar More<sup>3</sup>, Nandkishor Deshmukh<sup>4</sup>, Shankar Borkar<sup>5</sup>*

<sup>1</sup>Student of Bachelor of Pharmacy Shraddha Institute of pharmacy Kondala Zambre washim-444505

<sup>2,5</sup>Assistant Professor, Department of Quality Assurances Shraddha Institute of pharmacy Kondala Zambre washim-444505

<sup>3,4</sup>Assistant Professor, Department of Pharmaceutics Shraddha Institute of pharmacy Kondala Zambre washim-444505  
Shraddha Institute of pharmacy Kondala Zambre washim-444505

### ABSTRACT:

The Principle of UV-Visible Spectroscopy is based on the absorption of ultraviolet light or visible light by chemical compounds, which results in the production of distinct spectra. Spectroscopy is based on the interaction between light and matter one of the earliest instrumental techniques for analysis is UV-VIS spectroscopy. Many different types of materials can be characterized using UV-Vis spectroscopy. The UV-Vis delivers details based on the degree of absorption or transmittance of a varied wavelength of beam light and the various responses of samples. Radiant energy absorption by materials can be quantitatively described using the general law known as Beer's law. The UV-VIS spectrometer is simple to use and handle. Both qualitative and quantitative analyses can make use of it. The metal and metal oxide nanoparticles are typically characterized using wavelengths between 200 and 700 nm. The intricate mechanism of complexation between templates, monomer, and cross-linker during polymerization can also be better understood with the aid of the UV/Vis spectrum. It is quick, simple, and affordable characterization method. The composition and structure of the materials can be examined using the spectrum. These conclusions have uses in academia, business, medical labs, and chemical examination of environmental samples.

Rapid and easy analytical methods are needed due to increasing number of multicomponent formulations, biotherapeutic products and samples of complex matrix in que. Number of Ultraviolet (UV) spectrophotometric methods used for these purpose. Different types of UV spectrometric methods developed on the basis of principle of additivity, absorbance difference, processing absorption spectra. The aim of this review is to present information on simultaneous equation method, difference spectrophotometry, derivative spectrophotometry, absorbance ratio spectra, derivative ratio spectra, successive ratio - derivative spectra, Q-absorbance ratio method, absorptivity factor method, dual wavelength method, absorption factor method, multivariate chemometric methods, and isosbestic point method. A brief summary on theories, mathematical background and some applications of these methods are presented here

**Keywords:** Derivative spectrophotometry, Derivative ratio spectra, Isosbestic point UV-Visible Spectroscopy, UV-VIS Spectrometer, UV-VIS Spectrum method, Multivariate chemometric methods Ultraviolet spectroscopy, Simultaneous equation method

### INTRODUCTION TO SPECTROSCOPY:

Spectroscopy as a science began with Isaac Newton splitting light with a prism and was called optics. Therefore, it was originally the study of visible light which we call color that later under the studies of James Clerk Maxwell came to include the entire electromagnetic spectrum. Spectroscopy is the branch of science dealing with the study of interaction of electromagnetic radiation with matter. The most important consequence of such interaction is that energy is absorbed or emitted by the matter in discrete amounts called quanta. The absorption or emission processes are known throughout the electromagnetic spectrum ranging from the gamma region (nuclear resonance absorption or the Mossbauer effect) to the radio region (nuclear magnetic resonance). When the measurement of radiation frequency is done experimentally, it gives a value for the change of energy involved and from this one may draw the conclusion about the set of possible discrete energy levels of the matter. The ways in which the measurements of radiation frequency (emitted or absorbed) are made experimentally and the energy levels deduced from these comprise the practice of spectroscopy.

In routine practice, analyst has to perform rapid analysis of multicomponent formulations, biotherapeutic products and samples of a complex matrix. Number of ultraviolet (UV) spectrophotometric methods used for these purpose. However, among all of these methods, UV spectrophotometry is favorite tool. The basic principle behind the UV spectroscopy is absorption of visible and UV radiation (200– 400 nm) is associated with excitation of electrons, in both atoms and molecules, from lower to higher energy levels. Since the energy levels of matter are quantized, only light with the precise amount of energy can cause transitions from one level to another will be absorbed [1]. UV spectrophotometric methods based on principle of additivity and absorbance, recording and mathematical processing absorption spectra of standard solutions and sample solutions in same way or differently.

---

**TYPES OF UV SPECTROSCOPIC ANALYTICAL TECHNIQUES :**
**Simultaneous equation methods**

1. Derivative spectrophotometry
2. Absorbance ratio spectra
3. Derivative ratio spectra
4. Successive ratio - derivative spectra,
5. Q-absorbance ratio method
6. Absorptivity factor method
7. Dual wavelength method
8. Absorption factor method
9. Multivariate chemometric method
10. Isosbestic point method.

Spectroscopy is one of the most powerful tools available for the study of atomic and molecular structure and is used in the analysis of a wide range of samples. The study of spectroscopy can be carried out under following heads

**1) Atomic spectroscopy** : This deals with the interaction of electromagnetic radiations with atom which are most commonly in their lowest energy state called the ground state. The electronic absorption of electromagnetic radiation can occur only if the photon has an energy which is equal to the energy difference between two quantized energy levels, i.e.,  $\Delta E = h \nu$  where  $\Delta E$  is the energy difference between two quantum levels and  $\nu$  is the frequency of photon which can result in the electronic excitation. Applications of electronic spectroscopy in the field of chemistry are few. However, its importance has increased by the development of lasers.

**2) Molecular Spectroscopy** : This deals with the interaction of electromagnetic radiation with molecules. This results in transitions between rotational and vibrational energy levels in addition to electronic transitions. As a result, the spectra of molecules are much more complicated than those of atoms. Molecular spectra extend from the visible through infrared into the microwave region. Current interest in molecular spectroscopy is very great because the number of known molecules are extremely large as compared with free atoms. The various types of spectra given by molecular species, the regions in which these spectra lie and the energy changes that takes place in the molecules on absorption of radiation, are listed below.

a) Rotational (Microwave) Spectra These spectra arise due to transitions between the rotational energy levels of a gaseous molecule on the absorption of radiations falling in the microwave region These spectra are shown by molecules which possess a permanent dipole moment, e.g., HCl, CO, HO vapor, NO, etc.

b) Vibrational and Vibrational-Rotational (Infrared) Spectra These spectra arise due to transitions induced between the vibrational energy levels of a molecule on the absorption of radiation belonging to the infrared region. IR spectra are shown by molecules when vibrational motion is accompanied by a change in the dipole moment of the molecule. These spectra occur in the spectral range of 500-4000  $\text{cm}^{-1}$ .

c) Raman Spectra Raman spectra is related to vibrational and/or rotational transitions in molecules but in a different manner. In this case, what we measure is the scattering and not the absorption of radiation. An intense beam of monochromatic radiation in the visible region is allowed to fall on a sample and the intensity of scattered light is observed at right angles to the incident beam. Most of the scattered light has the same frequency as the incident beam (this is called Rayleigh scattering). Raman spectra are observed in the visible region, viz., 12,500-25,000  $\text{cm}^{-1}$ .

d) Electronic Spectra Electronic spectra arise due to electronic transitions in a molecule b absorption of radiations falling in the visible and ultraviolet regions. While electronic spectra in the visible region span 12,500-25,000  $\text{cm}^{-1}$ , those in the ultraviolet region span 25,000-70,000  $\text{cm}^{-1}$ . As electronic transitions in a molecule are invariably accompanied by vibrational and rotational transition the electronic spectra of molecules are highly complex.

e) Nuclear magnetic resonance NMR spectra arise due to transitions induced between the nuclear spin energy levels of a molecule in an applied magnetic field. NQR spectra arise due to the transitions between the nuclear spin energy of a molecule arising from the interaction of the unsymmetrical charge distribution in nuclei with the electric field gradients (EFG) which arise from the bonding and non-bonding electrons in the molecule, NMR and NQR spectra span the radiofrequency regions, viz., 5-100 MHz.

f) Mossbauer Spectra Also called Nuclear Gamma Resonance Fluorescence (NRF) Spectra Mossbauer spectra is a type of nuclear magnetic resonance spectra. However, while NMR spectra arise due to absorption of low energy photons of frequency around 60 MHz, Mossbauer spectra result from absorption of high energy  $\gamma$ -photons of frequency around 10<sup>13</sup> MHz by the nuclei. Gamma ray spectra have been used specifically for the study of compounds of iron and tin. In this case,  $\gamma$ -radiations from <sup>57</sup>Co-source are allowed to fall on a sample in which the iron nuclei are in an environment identical with that of the source atoms. This gives rise to the resonant absorption of  $\gamma$  rays. The splitting in Mossbauer lines are found to be of the same order as in NMR spectroscopy.

a) Rotational (Microwave) Spectra These spectra arise due to transitions between the rotational energy levels of a gaseous molecule on the absorption of radiations falling in the microwave region. These spectra are shown by molecules which possess a permanent dipole moment, e.g., HCl, CO, HO vapor, NO, etc. b) Vibrational and Vibrational-Rotational (Infrared) Spectra These spectra arise due to transitions induced between the vibrational energy levels of a molecule on the absorption of radiation belonging to the infrared region. IR spectra are shown by molecules when vibrational motion is accompanied by a change in the dipole moment of the molecule. These spectra occur in the spectral range of 500–4000  $\text{cm}^{-1}$ . c) Raman Spectra Raman spectra is related to vibrational and/or rotational transitions in molecules but in a different manner. In this case, what we measure is the scattering and not the absorption of radiation. An intense beam of monochromatic radiation in the visible region is allowed to fall on a sample and the intensity of scattered light is observed at right angles to the incident beam. Most of the scattered light has the same frequency as the incident beam (this is called Rayleigh scattering). Raman spectra are observed in the visible region, viz., 12,500–25,000  $\text{cm}^{-1}$ . d) Electronic Spectra Electronic spectra arise due to electronic transitions in a molecule on absorption of radiations falling in the visible and ultraviolet regions. While electronic spectra in the visible region span 12,500–25,000  $\text{cm}^{-1}$ , those in the ultraviolet region span 25,000–70,000  $\text{cm}^{-1}$ . As electronic transitions in a molecule are invariably accompanied by vibrational and rotational transition the electronic spectra of molecules are highly complex. e) Nuclear magnetic resonance NMR spectra arise due to transitions induced between the nuclear spin energy levels of a molecule in an applied magnetic field. NQR spectra arise due to the transitions between the nuclear spin energy of a molecule arising from the interaction of the unsymmetrical charge distribution in nuclei with the electric field gradients (EFG) which arise from the bonding and non-bonding electrons in the molecule, NMR and NQR spectra span the radiofrequency regions, viz., 5–100 MHz. f) Mossbauer Spectra Also called Nuclear Gamma Resonance Fluorescence (NRF) Spectra Mossbauer spectra is a type of nuclear magnetic resonance spectra. However, while NMR spectra arise due to absorption of low energy photons of frequency around 60 MHz, Mossbauer spectra result from absorption of high energy  $\gamma$ -photons of frequency around 10<sup>13</sup> MHz by the nuclei. Gamma ray spectra have been used specifically for the study of compounds of iron and tin. In this case,  $\gamma$ -radiations from <sup>57</sup>Co-source are allowed to fall on a sample in which the iron nuclei are in an environment identical with that of the source atoms. This gives rise to the resonant absorption of  $\gamma$ -rays. The splitting in Mossbauer lines are found to be of the same order as in NMR spectroscopy.

## UV SPECTROSCOPY:

UV-Vis (ultraviolet visible) spectroscopy is frequently used to provide characterization data for a variety of materials. Inorganic or organic, solid or liquid groups, such as organic molecules and functional groups, can be observed using UV-Visible spectroscopy, as can reflectance measurements for coatings, paints, textiles, biochemical analysis, dissolution kinetics, band gap measurements, etc. Depending on the degree of absorbance or transmittance of a different wavelength of beam light and the various responses of samples, the UV-Vis provides these details.

### *Why is UV spectroscopy used?*

UV spectroscopy is useful for investigating the electronic structures of unsaturated molecules and for measuring the extent of their conjugation.

### *Electromagnetic Spectrum:*

The ability of electromagnetic radiation to discretely interact with atoms and molecules and produce distinctive absorption or emission profiles is essential for spectroscopic activities. The wavelength of electromagnetic radiation is the characteristic that governs the perceived color spectrum. The visible section of the electromagnetic spectrum is that portion of the spectrum that the human eye can see. These visible wavelengths span a region between 400 and 800 nm. A specific wavelength or color of visible light corresponds to the optical density when it is measured with spectrophotometers. This light is absorbed, vanishes, and becomes invisible. The approximate complimentary connection between the light wavelengths that are absorbed and those that are transmitted is shown in Figure. For instance, the complimentary color of light, orange, would be strongly absorbed by a blue substance.

The distance between neighbouring peaks or troughs is known as the wavelength. The following straightforward equation can be used to define the wavelength, of EMR as a function of its frequency  $\nu$ , and the speed of light

$$c = \nu \lambda \quad (1)$$

EMR has both particle and wave behaviour (the dual nature of light), and the relationship between energy and the wavelength of such a particle, a photon, is given by the equation

$$E = hc / \lambda \quad (2)$$

Where  $c$  is the speed of light in a vacuum ( $2.998 \times 10^8 \text{ ms}^{-1}$ ),  $h$  is the Planck's constant ( $6.63 \times 10^{-34} \text{ Js}$ ),  $E$  is the photon's energy, and  $\lambda$  is the wavelength in nm.

### *Beer's Law:*

The intensity of the incident radiation ( $I_0$ ) will be higher than the emerging radiation when a beam of electromagnetic radiation passes through an absorbing material ( $I$ ). The general rule known as Beer's law can be used to quantitatively describe how radiant energy is absorbed by materials. According to Beer's law, the amount of radiation that is absorbed (absorbance,  $A$ ) or transmitted by a solution or medium is inversely related to the amount of the

absorbing substance that is present,  $c$  (moles per liter), and the length of the radiation's passage through the sample,  $b$ , (centimeters). Therefore, a plot of absorbance against concentration should result in a line that is straight and has a slope equal to  $\epsilon b$ , passing through the origin.

$A = -\log(I/I_0) = \epsilon bc$  where  $\epsilon = k/2.303$  The molar absorptivity is a constant  $\epsilon$  that is unaffected by concentration or path length. If the route length and molar absorptivity are known, Beer's law equation can be used to calculate the concentration of an organic molecule by locating its highest absorbance in the UV-Vis absorption spectrum.

---

### Principle of Uv-Vis Spectroscopy:

When radiation induces an electronic transition in a molecule or ion's structure, the object will exhibit absorption in the visible or ultraviolet range. As a result, when a sample absorbs light in the ultraviolet or visible range, the molecules inside the sample experience a change in their electronic state. Electrons will be promoted from their ground state orbital to a higher energy, excited state orbital by the energy from the light, or anti-bonding orbital. Potentially, three types of ground state orbitals may be involved.

- i.  $\sigma$  (bonding) molecular
- ii.  $\pi$  (bonding) molecular orbital
- iii.  $n$  (bonding) atomic orbital

In addition, two types of anti-bonding orbitals may be involved in the transition

- i.  $\sigma^*$  (sigma star) orbital
- ii.  $\pi^*$  (pi star) orbital

There is no such thing as an  $n^*$  anti-bonding orbital as the  $n$  electrons do not form bonds). Thus, the following electronic transitions can occur by the absorption of ultraviolet and visible light.

- i.  $\sigma$  to  $\sigma^*$
- ii.  $n$  to  $\sigma^*$
- iii.  $n$  to  $\pi^*$
- iv.  $\pi$  to  $\pi^*$

Due to their high energy requirements, the  $\sigma$  to  $\sigma^*$  and  $n$  to  $\sigma^*$  transitions both take place in the far ultraviolet area or sporadically in the range of 180–240 nm. Saturated groups consequently do not show high absorption in the common UV range. In contrast to transitions to the  $\pi^*$  anti-bonding orbital, transitions from then to the  $\pi^*$  and to  $\pi^*$  type occur in molecules with unsaturated centers. They need less energy and take place at longer wavelengths. It will soon be clear that molecule structure controls both the absorption's maximum wavelength and its intensity. If a molecule's chemical structure is changed, transitions to the  $\pi^*$  anti-bonding orbital that normally take place in the UV range could very well occur in the visible region. Many inorganic compounds in solution also show absorption in the visible region. These include salts of elements with incomplete inner electron shells (mainly transition metals) whose ions are complexed by hydration. Such absorptions arise from a charge transfer process, where electrons are moved from one part of the system to another by the energy provided by the visible light

---

### Instrumentation:

The Essential components of UV-VIS Spectrophotometer are as follows:

1. Sources (UV and visible)
2. Monochromator
3. Sample containers (Cuvette)
4. Detector
5. Amplifier and recorder

1. SOURCES: A continuous source, or one that produces radiation at a variety of wavelengths, is necessary for UV-Vis Spectroscopy. Assorted UV radiation sources include the following:

- 1) Hydrogen lamp: Hydrogen lamps are reliable, steady, and continuously emit radiation between 160 and 380 nm. It consists of hydrogen gas at high pressure, which causes an electrical discharge. The excited hydrogen molecules produce radiation.
- 2) Deuterium lamp: A gas discharge lamp called a deuterium lamp is frequently employed as a UV source. It emits radiation in the 160–450nm range. It costs more than a hydrogen lamp.

- 3) 3) Tungsten lamp: The most typical light source utilized in spectrophotometers is the tungsten lamp. With a wavelength range of roughly 330 to 900 nm, it comprises of a tungsten filament encased in a glass envelope and is utilized for the visible spectrum.
- 4) 4) Xenon discharge lamp: A xenon lamp is a discharge light source that contains xenon gas inside a bulb. Radiation from xenon ranges from 250 to 600 nm

2. MONOCHROMATOR: By filtering out undesirable wavelengths from the radiation source light, a monochromator creates monochromatic light. Through the entrance slit, multi-wavelength polychromatic light enters the monochromator. Following collimation, the beam is directed at an angle toward the dispersion component. The grating or prism separates the beam's wavelengths into their individual components. Only radiation of a specific wavelength exits the monochromator through the exit slit when the dispersing element or the exit slit are moved.

Types of monochromators:

- 1) Prism monochromator
- 2) Grating monochromator

3. SAMPLE CONTAINERS (CUVETTE): Cuvettes are sample containers that are transparent to all wavelengths of light flowing through them and are used to hold samples for spectroscopic measurements. The cuvette is composed of quartz, is square in shape, has a 1 cm route length, and may be utilized for wavelengths between 190 and 200 nm.

4. DETECTORS: Light energy is converted by detectors into electrical impulses that are read out by readout devices. The transmitted radiation strikes the detector, determining the amount of radiation absorbed by the sample. The absorption spectrophotometer's apparatus uses the following types of detectors.

Types of Detectors:

- i. Barrier layer cell/Photovoltaic cell
- ii. Phototubes/ Photo emissive tube
- iii. Photomultiplier tube

---

### **Different UV spectrophotometric multicomponent analysis methods :**

There are some methods

1. Simultaneous equation method: If a sample contains two absorbing drugs (x and y) each of which absorbs at the  $\lambda_{\max}$  of the other, it may be possible to determine both drugs by the technique of simultaneous equation (Vierordt's method) provided that certain criteria apply. The information required is

- The absorptivities of x at  $\lambda_1$  and  $\lambda_2$ ,  $a_{x1}$  and  $a_{x2}$  respectively
- The absorptivities of y at  $\lambda_1$  and  $\lambda_2$ ,  $a_{y1}$  and  $a_{y2}$  respectively
- The absorbance of the diluted samples at  $\lambda_1$  and  $\lambda_2$ ,  $A_1$  and  $A_2$  respectively.

Let  $C_x$  and  $C_y$  be the concentration of x and y respectively in the diluted samples. Two equations are constructed based upon the fact that at  $\lambda_1$ , the absorbance of the mixture is the sum of the individual absorbance of x and y.

#### ***Difference spectrophotometry:***

The selectivity and accuracy of spectrophotometric analysis of samples containing absorbing interferences may be markedly improved by the technique of difference spectrophotometry. The essential feature of this method is that the measured value is the absorbance difference ( $\Delta A$ ) between two equimolar solutions of the analyte in different chemical forms which exhibit different spectral characteristics.

#### ***Derivative spectrophotometry (DS):***

DS involves the conversion of a normal spectrum (fundamental, zero-order spectrum) to its first, second or higher derivative spectra by differentiating absorbance of the sample with respect to wavelength ( $\lambda$ ). The differentiation of zero-order spectrum can lead to separation of overlapped signals, elimination of background caused by presence of other compounds in a sample, improvement of resolution of mixtures as it enhances the detectability of minor spectral features, and enhancement of sensitivity and specificity.

**Derivative ratio spectra method:**

This simple spectrophotometric method, developed by Salinas et al. is based on the derivation of the ratio spectra for resolving binary mixtures. It permits the use of the wavelength of highest value of analytical signals with several maxima and minima, which give an opportunity for the determination of active compounds in the presence of other compounds and excipients which could possibly interfere in the assay

**CONCLUSION:**

If used with the right standard curve and applied to pure substances, UV-visible spectroscopy is a reliable, straightforward, and affordable approach for estimating the concentration of absorbing species. One of the crucial methods for analyzing the optical characteristics of PMCs is UV-Vis spectroscopy. It clarifies the relationship between the matrix and the nanofiller and examines how the nanofillers contribute to the enhancement of the properties of the nanocomposites. To assess the intended optical properties of nanofillers in a polymer matrix, UV-Vis spectroscopy is a crucial technique. The polymer nanocomposites with some optically responsive nanofillers, such as metals, semiconductor nanocrystals, and nano oxides, are characterized using the UV-Vis spectroscopic approach in order to produce functional materials with significant technological applications. UV-Vis Spectroscopy (or Spectrophotometry) is a quantitative technique used to measure how much a chemical substance absorbs light.

**REFERENCES:**

1. Gurdeep R. Chatwal, Anand Sharma Instrumental method of Analysis 5th edition, Himalaya publishing house PVT.LTD, 2002 pp 2.566-2.587
  2. J L Aldabib, M F Edbeib, "The Effects of Concentration based on the absorbance form the Ultraviolet Visible Spectroscopy", International Journal of Science Letters, 2020, 2(1): 1-11
  3. G Verma, DR. M Mishra, "Development and Optimization of UV VIS Spectroscopy-A Review", World Journal of Pharmaceutical Research, 2018,7(11),1170-1180
  4. Ganesh S, et al. "A Review on Advances in UV Spectroscopy", Research J. Science and Tech. 2020,12(1),47- 51
  5. Amira H. Kamal, Samah F, et al. "A REVIEW ON UV SPECTROPHOTOMETRIC METHODS FOR SIMULTANEOUS MULTICOMPONENT ANALYSIS", European Journal of Pharmaceutical and Medical Research, 2016,3(2), 348-360
  6. M. Patil, "Advantages and Disadvantages of UV Visible Spectroscopy", NOV 2018, <https://chrominfo.blogspot.com/2018/11/advantages-and-disadvantages-of-uv.html?m=1>
  7. N Lavanya, et al. "Simultaneous Determination and Comparison of Sun Protection Factor of Various merchandised sunscreen formulations by using UV Spectrophotometer", Asian Journal of Pharmaceutical Analysis, 2022,12(2),111-114
  8. Priya, Archana G. "UV Spectrophotometric Method Development and Validation for Amlodipine Besylate in Bulk and Tablet Dosage Form", Asian Journal of Pharmaceutical Analysis, 2022,12(2),94-98
  9. Maruthi R, et al. "Analytical Method Development and Validation of Teneigliptin by UV Spectroscopy", Research Journal Pharm and Tech, 2021,14(1),75-78
  10. Kosyan A, Sytar O, "Implications of Fagopyrin Formation In Vitro by UV Spectroscopic Analysis", Molecules 2021,26,2013,1-10
  11. Sandro Jurinovich, Valentina D. "Digital Tool for the Analysis of UV-Vis Spectra of Olive Oils and Educational Activities with High School and Undergraduate Students", Journal of Chemical Education 2022, 99,787-798
  12. Hajian R, Shams N, Kaedi I, Club YR. Application of ratio derivative spectrophotometry for simultaneous determination of naphazoline and antazoline in eye drops. E J Chem 2010;7:1530-8.
  13. Fayez YM, Elghobashy MR, Goda ZM, Shehata MA. Comparative study on four spectrophotometric methods manipulating ratio spectra for the simultaneous determination of binary mixture of diflucortolone valerate and isoconazole nitrate. Bull Fac Pharm 2016;54:2-9.
- Pravinbhai P, Patel PB. Development and validation of ratio-derivative spectrophotometric method for simultaneous estimation of gabapentin, methylcobalamin and alpha lipoic acid in tablet formulation. J Pharm Res 2013;6:609-14.
1. 36. Afkami A, Bahram M. Successive ratio-derivative spectra as a new spectrophotometric method for the analysis of ternary mixtures for the analysis of ternary mixtures. Spectrochim Acta Part A 2005;61:869-77.
  2. 37. Abdelwahab NS. Spectrophotometric methods for simultaneous determination of carvedilol and hydrochlorothiazide in combined dosage form. Arab J Chem 2016;9:S355-60.
  3. 38. Patil PA, Raj HA, Sonara GB. Q-absorbance ratio spectrophotometric method for simultaneous determination of atenolol and ivabradine hydrochloride in synthetic mixture. Pharm Biol 2016;3:224-30.

- 
4. 39. Patil A, Surana S. Q-absorbance and multicomponent UV spectrophotometric methods for simultaneous estimation of propranolol. Scholars research library. Pharm Chem 2016;3:404-8.
  5. 40. Singh G, Kumar D, Sharma D, Singh M. Q-absorbance ratio spectrophotometric method for the simultaneous estimation of prednisolone and 5-amino salicylic acid in tablet dosage form. J Appl Pharm Sci 2012;2:222-6.
  6. Patel A, Firke SD, Bari SB, Ranoliya JR. Development and validation of UV-spectrophotometric method for simultaneous estimation of naproxen and paracetamol by Q-absorbance ratio method. Int J Pharm Res Allied Sci 2014;3:57-63.
  7. Pandey G, Mishra B. A new analytical Q-absorbance ratio method development and validation for simultaneous estimation of lamivudine and isoniazid. ISRN Spectrosc 2013;2013:1-5.
  8. Patel P, Patel D, Desai S, Meshram D. Development and validation of analytical methods for simultaneous estimation of difluprednate and gatifloxacin in ophthalmic emulsion by Uv-visible spectroscopy. Int J Pharm Sci 2014;3:1-10.
  9. Samir A, Salem H, Abdelkawy M. New developed spectrophotometric method for simultaneous determination of salmeterol xinafoate and fluticasone propionate in bulk powder and seritide ödiskus inhalation. Bull Fac Pharm Cairo Univ 2012;50:121-6.
  10. Saleh SS. A comparative study of validated spectrophotometric and TLC-spectrodensitometric methods for the determination of sodium cromoglicate and fluorometholone in ophthalmic solution. Saudi Pharm J 2013;21:411-21.
  11. Sharma R, Mehata K. Simultaneous spectrophotometric estimation of tenofovir disoproxil fumarate and lamivudine in three component tablet formulation containing efavirenz. Indian J Pharm Sci 2010;74:527-30