

International Journal of Research Publication and Reviews

Journal homepage: www.ijrpr.com ISSN 2582-7421

Smart Textile for Healthcare Monitoring

Ravi Kumar M¹, Jayaprakash S², Racliana NS ³, Lekha H⁴, Priya S⁵ and Rashmitha Reddy R⁶

Assistant Professor, Cambridge Institute of Technology Bengaluru, India Student, Cambridge Institute of Technology Bengaluru, India DOI: <u>https://doi.org/10.55248/gengpi.5.0424.1006</u>

ABSTRACT

An electrocardiography SoC, flexible electrodes, battery, and antenna are all combined into a form- fitting cloth. This "smart shirt" records a 12-lead ECG, which is the clinical standard. Using a flexible antenna and on-chip ISM band radio, the data is securely transferred wirelessly, consuming less than ImW and enabling secure, continuous cardiac monitoring on a smartphone.

Keywords: Electrocardiography (ECG), System-on-Chip (SoC), Flexible electrodes, Battery, 12-lead ECG, Wireless data transfer, ISM band radio

Introduction

Technological breakthroughs have caused a tremendous upheaval in the healthcare sector in recent years. Smart textiles are one of these inventions that really sticks out as a ground-breaking idea that might completely change healthcare monitoring. Smart textiles provide a unique, non-invasive, and extremely successful method to patient care by seamlessly integrating electronics into fabrics. Conventional healthcare monitoring techniques frequently entail large, unpleasant gadgets that limit patients' range of motion and jeopardize their comfort. Smart textiles, on the other hand, do away with these limitations by integrating sensors, actuators, and other electrical parts right into garments. This allows for ongoing physiological parameter and vital sign monitoring without interfering with the patient's day-to-day activities. Smart textiles have a broad range of uses in healthcare that go beyond simple monitoring. These textiles provide a comprehensive solution for proactive healthcare management, ranging from tracking physical activity and identifying irregular heart rhythms to monitoring blood glucose levels and evaluating sleep quality. Moreover, real-time data transmission and analysis are made easier by their seamless integration with wearable technology and wireless communication systems, which allows for prompt intervention and individualized treatment plans. Improving patient outcomes and increasing the effectiveness of healthcare delivery are two major benefits of using smart textiles in healthcare monitoring. Smart textiles enable medical practitioners to make well-informed judgments and take preemptive measures to avert unfavorable outcomes by giving them access to precise, real-time data. As sensor technology, materials science, and data analytics continue to progress, smart textiles have the potential to become a vital component of the healthcare system. Smart textiles have the potential to revolutionize healthcare delivery by enabling continuous, non-invasive monitoring of patients' health state. This rrnght lead to a more proactive, tailored, and patient-centric approach to healthcare delivery. Smart textiles are expected to be crucial in creating a healthier and more connected world as we move forward with health care.





Fig: Block diagram of Smart Textile For Healthcare Monitoring

- 1. Smart Textile Fabric: Fabric with conductive and flexible materials that is integrated with Textronic E-Textile Sensors. records physiological signals from the body, such as the ECG.
- 2. ECG **Sensor:** Contains an analog front-end, a signal amplifier, and sample electrodes. Captures and enhances the electrical impulses that the body produces.
- Analog Processing: Uses a variety of filters for signal conditioning as well as an analog-to-digital converter (ADC). transforms analog ECG sensor signals into digital format and prepares the signal for additional processing.
- 4. Arduino Nano Microcontroller: This microcontroller handles communication, data processing, and component interface enables communication, applies methods for data compression and heart rate analysis, and processes digital information.
- 5. Digital Signal Processing: Performs data compression, transmission, and heart rate analysis among other digital signal processing activities.evaluates ECG readings, calculates heart rate, handles data compression, and controls communication with other devices.
- 6. Display Interface: A visual output interface for information display (LCD, LED, etc.). gives the user access to pertinent data, such heart rate or system condition.

Proposed System:

The utilization of smart textiles for ECG measurements is a revolutionary development in the wearable health technology space. This suggested approach aims to establish a network of discrete yet extremely effective ECG electrodes by sewing conductive textiles or threads into garments. A small, light monitoring device that can be discretely attached or integrated in the textile is a crucial part of this system. The smart textile is made of materials that favor durability and breathability, with an emphasis on user comfort. Because washable materials are used, the item is practical and appropriate for daily usage. This smart textile technology has applications in a range of healthcare settings, including as preventive care and remote patient monitoring. Healthcare practitioners can evaluate and intervene remotely in the event of an abrupt cardiac incident thanks to continuous ECG monitoring. Thus, the suggested solution solves the demand for inconspicuous, ongoing, and easily accessible ECG monitoring by fusing technological innovation with user-centric design..

Literature Survey

1. Authors: Mathew Ridder and Luis Lopez Ruiz.

TITLE: E-textiles and Multi-i Model Harvesting for Self-Powered Cardiac Monitoring and Vigilance Maintenance, August 23, 2001.

The nexus between engineering and medicine has given rise to remote patient monitoring. Many of these applications fall under the category of cardiac monitoring, and the use of wearable monitors in routine clinical practice has been driven by the necessity to record fleeting cardiac events. The problems

of user comfort, skin irritation, and motion artifacts are addressed by a customized ECG shirt consisting of knitted compression fabric with embedded dry electrodes.

2. Hui Yang, Chen Kan, and Wensin Tong.

TITLE: Wearable Textile Sensitivity Analysis for ECG Sensing.

Biomedical & Health Informatics 2018 IEEE EMBS International Conference.

The new generation of wearable electrocardiogram (ECG) sensing equipment is brought about by rapid advancements in material science and mobile technology. Specifically, because of their great reusability and flexibility, sensing textiles have found extensive application in cardiac monitoring. The sensitivity of textile-based ECG sensing to four factors—contact pressure, textile positioning, user activity, and muscle activity—is experimentally investigated in this work.

3. AUTHORS: V S Prakash, Sarthak Sahu, Avvaru Srinivasulu, and N. Sriram.

TITLE: Smart Testile Electrode Belt For Recordings - A Pilot Study with Indian Population. The International Conference on Signal Processing and Communication (ICSPC - 2019) is taking place in 2019.

This work demonstrates that the suggested design of a silver-plated nylon woven (SiPNyW) textile electrode can record an ECG signal while the patient is immobile. This signal quality is comparable to that obtained by a typical Ag/AgC1 ge1 electrode.ECGs were recorded using a textile electrode belt with a Lead-I configuration based on SiPNyW. For the purpose of HRV analysis, 147 healthy Indian volunteers between the ages of 18 and 45 participated in this study. According to the findings, the quality of the ECG signal obtained by cloth electrodes and Ag/AgC1 electrodes was about equal.

4. DR. Malti Bansal, Bani Gandhi is the author.

TITLE: Smart ECG Monitoring Using Textiles Based on CNTs.

The third conference on recent developments in power engineering, automation, and control is scheduled for 2019.

A key technological advancement in the healthcare industry is the evolution of flexible, conformal, and stretchy electronics integrated with commercially accessible electronic modules placed on a soft and elastic substrate. The needs for healthcare have changed over time, shifting from being hospital-centric to home-centric. ECG signals from the patient, which can subsequently be sent to the physician for professional guidance. Thus, we shed light on CNT-based textiles for intelligent ECG monitoring in this study.

5. Authors: Manabu Yoshida and Toshihiro Takeshita.

Title: CUBIC Embedding Amplifier Circuit with Flocked Electrode for Smart ECG Testile

Application.2010; Tsukuba, National Institute of Advanced Industrial Science and Technology

They presented a unique dry electrode configuration for the development of smart electrocardiogram (ECG) textile. Ag-plated fiber was flocking onto a dimethylpolysiloxane (PDMS) cube using electrostatic flocking technology as the dry electrode. Its construction allowed for consistent contact pressure. It is possible to create a highly functional and versatile vital signs monitoring system by employing this electrode structure to integrate electrical devices like a MEMS sensor, battery, MCU, and RF-IC inside the cubic dry electrode.

6. AUTHORS: Tsair Kao and Chien-Lung Shen.

TITLE: Wearable Band Monitoring Exercise ECG Using a Fabric-Based Sensor,2006

Smart textiles have garnered a lot of attention lately, particularly for wearable clothing that can capture vital signs. Our study presents a system for monitoring the exercise electrocardiogram (ECG) using fabric-based garment sensing; a Bluetooth-based wireless communication configuration for transferring the signals into a PDA; and an algorithm for extracting the appropriate ECG signals from the raw data. The testing findings while at rest, walking, or running demonstrate that this wearable clothing system can accurately and steadily obtain the RRI and ECG.

7. AUTHORS: Joshua Di Tocco, Carlo Massaroni, and Daniella Lo Presti.

TITLE: Sensor positioning affects cardiac monitoring using a smart textile based on polymer-

encapsulated FBG,2018 is the year.

One of the most cutting-edge categories of wearables for continuous, non-invasive heart activity monitoring is smart textiles. One well-known method relies on identifying the vibrations that the heartbeat causes on the surface of the chest. Various sensor placements have been studied in the literature, but it doesn't seem like there are any widely acknowledged standard points for the identification of heart-induced motions.

8. Writers: W. Chen, S. Bangag Oetomo, M.D.

Title: Developing Testile Neonatal ECG Monitoring Capabilities Through Multi-Sensor Captures, August 30, 2011, Boston, Massachusetts, USA, 33rd Annual International Conference of the IEEE EMBS conference.

Because textile electrodes can create inadvertent low signal quality and motion artifacts, it might be difficult to maintain accurate monitoring when building an ECG monitoring system incorporated with textile electrodes for comfort. We suggest the ideas of "diversity measurement" and "context awareness" to enhance the design of a cozy monitoring system for prematurely born infants in the Neonatal Intensive Care Unit (NICU) and to increase dependability. We have developed design criteria and meta-insights regarding the function of clinical validation in the design process as a result of our exploratory system level approach.

9. Hassan Bajwa, Linfeng Zhang, and Ali Alzaidi are the authors. TITLE: Wireless ECG System Based on Smart Textiles, 2011 is the year.

There have also been new reports on non-contact electrodes and smart textiles. This research presents a wireless ECG system that makes use of smart textile sensors. An essential component of a wireless ECG system is a non-contact ECG electrode that was created using clever materials and several textile technologies. Together with the system level radio design that might be merged with the suggested ECG electrode and textile-based conformable antenna, this study will also provide the design of the integrated conformal patch antenna based on textiles.

10. Authors: Bani Gandhi, Malti Bansal.

TITLE: ECG Electrodes for Smart ECG Monitoring for biomedical applications. Third International Conference on I-SMAC (Internet of Things in Social Media, Mobile, Analytics, and Cloud) Proceedings (I-SMAC 2019).

The healthcare sector is under tremendous pressure due to the world's population growth. In addition to being more linked and portable, healthcare systems should be dependable. The sensors that are utilized to measure the Electrocardiogram (ECG) for different CVDs are the most important thing. Carbon nanotubes (CNTs) have several benefits, including flexibility, conductivity, and low weight. As a result, they can be incorporated into fabrics or utilized to create electrodes. Therefore, those who have CVDs can use these electrodes or sensors in their daily lives to help detect the condition early and improve their quality of life.

Summary of Literature Review

TABLE: Literature review summary

Published	Author name	Title of the	Proposed technique	Limitations
year		paper		
2018	Hui Yang, Chen Kan, and Wenxin	Sensitivity Analysis Of Wearable Textiles	Bitalino Sensor	Cost of the Sensor is more
[1]	Tong.	For ECG Sensing.		
2001	Mathew Ridder and	Self Powered Cardiac	Wearable monitors. Self	Battery life lirrrit the duration.
[2]	Luis Lopez Ruiz.	Monitoring- Maintaining Vigilance With Mult- i Model Harvesting And E-Textiles.	powered system.	Represent high noise.
2019	VS Prakash, Sarthak	Smart Textile	Ag/AgCl gel electrode.	Gel may cause irritation.
[3]	Sahu, Avvaru Srinivasulu, and	Electrode Belt For Recordings	Conductive fabric. AD8232 single lead	
	N. Sriram.	- A Pilot Study with Indian Population.		
2019	Malti Bansal, Bani	CNT Based Textiles	MWCNT(Multi walled	Biocompatibility concerns.
[4]	Gandhi is the author.	For Smart ECG Monitoring.	nanotubes) SWCNT(Sing1e walled nanotubes)	Washability & durability.
2010	Manabu Yoshida	Cubic Flocked	Dry electrode known as	Electrical insulations
[5]	and Toshihiro	Electrode Embedding	dimethyl poly siloxane	Integration Challenges.
	i arcsinta.	Smart ECG Textile	MCU,RF—IC are	Biocompanointy.
		Application.	embedded in PDMS cube	

2006 [d]	Tsair Kao and Chien-Lung Shen.	Wearable Band Using a Fabric- Based Sensor for Exercise ECG Monitoring.	Fabric based garment sensing system	Potential interference from motion artifacts during physical exercise
2019 [7]	Joshua Di Tocco, Carlo Massaroni, and Daniella Lo Presti.	ECG Electrodes for Smart ECG Monitoring for biomedical applications	Integration of carbon nanotubes (CNTs) into wearable electrocardiogram (ECG) sensors or electrodes.	High sensitivity and specificity
2018 [8]	W. Chen, S. Bangag Oetomo, M.D.	Cardiac monitoring with a smart textile based on polymer- encapsulated FBG: influence of sensor positioning.	Computational Modeling	Chest surface vibrations can be affected by motion artifacts caused by factors such as body movement, respiration, and external environmental disturbances.
2011 [9]	Hassan Bajwa, Linfeng Zhang, and Ali Alzaidi are the authors.	Designing For Reliable Textile Neonatal ECG Monitoring Using Multi- Sensor Recordings.	System-Level Approach, Diversity Measurement	Complexity in hardware design
2011 [10]	Bani Gandhi, Malti Bansal.	Smart Textiles Based Wireless ECG System.	Textile—Based Conformal Patch Antenna	Signal Quality and Reliability

Conclusion and Future Scope

In order to advance the automotive and medical industries, smart textiles are used in a variety of industrial and healthcare sectors. Their monitoring as well as other technological features are crucial. Innovation and investment in smart textiles benefit each of the corresponding industries. A number of issues, including the weaving techniques and flexibility, need to be fixed. The smart textiles will be even more beneficial if the silica fiber is swapped out for another viable fiber. There is little doubt that smart textiles will make a big impact in the realms of technology and medicine.

References

- 1. Ankhili, A., Tao, X., Koncar, V.; Cochrane, C.; Coulon, D. Textile electrodes that are washable and dependable are integrated into undergarments to monitor electrocardiography (ECG). Materials 2018, 11, 256. [Scholar Google] [Cross Reference] [Version Green]
- Duan, Y.; Xu, F.; Huang, Y.; Wu, H.; Yin, Z. Energy Harvesters: Transitioning from Flexibility to Stretchability for Wearable and Stretchable Electronics. 201d, 28, 9881-9919; Adv. Mater. [Scholar Google] [Cross Reference]
- Piezoelectric Energy Harvesting from Intelligent Textiles by Waqar, S., Wang, L., and John, S. Elsevier, Amsterdam, The Netherlands, 2015, Electronic Textiles, pp. 173-197. 978-0-08-100201-8 is the ISBN. [Scholar Google]
- 4. Thermal Management of Wearable and Implantable Electronic Healthcare Devices: Perspective and Measurement Methodology, Bahru, R., Hamzah, A.A., Mohamed, M.A. Global Journal of Energy Research 2020, 45, 1517-1534. [Scholar Google] [Cross Reference]
- 5. Handbook of Flexible and Stretchable Electronics, M.M. Hussain and N. El-Atab, CRC Press, Boca Raton, FL, USA, 2019. [Scholar Google]
- Application Challenges in Fiber and Textile Electronics: Wang, L., Fu, X., He, J.; Shi, X.; Chen, T.; Chen, P.; Wang, B.; Peng, H. Adv. Mater. 2020, 32, 1901971. [Scholar Google] [Cross Reference] [PubMed]
- Joining Technologies for Electronic Textiles, Mecnika, V.; Scheulen, K.; Anderson, C.F.; Hörr, M.; Breckenfelder, C. Elsevier, Amsterdam, The Netherlands, 2015, pp. 133-153, in Electronic Textiles. 978-0-08-100201-8 is the ISBN. [Scholar Google]
- 8. uz Zaman, S.; Tao, X.; Cochrane, C.; Koncar, V. Understanding the Washing Damage to Textile ECG Dry Skin Electrodes, Embroidered and Fabric-Based; Set up of Equivalent Laboratory Tests. 2020, 20, 1272, Sensors. [Scholar Google] [Cross Reference] [PubMed] [Version Green]

- 9. Nigusse, A.B.; Langenhove, L.V.; Fante, K.A.; Malengier, B.; Tseghai, G.B. An overview of the integration of conductive materials with textile structures. 2020, 20, d910, Sensors. [Scholar Google] [Cross Reference] [PubMed]
- 10. Yang, Z., Yang, Y., and Ren, T.; Pang, Y. Wearable Electronics for Human Physiological Information Detection Using 2D Materials. 2020, 16; 1901124; small. [Scholar Google] [Cross Reference]
- 11. Towards Embroidered Circuit Board From Conductive Yarns for E-Textiles, Ismar, E., Tao, X., Rault, F., Dassonville, F., and Cochrane, C. IEEE Access 2020, 7, 9. [Scholar Google] [Cross Reference]
- 12. Let's Make Design Thinking Meaningful for Textiles: Valentine, L., Ballie, J., Bletcher, J., Robertson, S., and Stevenson, F. 2017, 20, S9d4-S976 in Design J. [Scholar Google] [Cross Reference]
- Scataglini, S., Gallant, J., and Andreoni, G. A Review on Smart Clothing in Military. published in WearSys '15: Proceedings of the Workshop on Wearable Systems and Applications, Florence, Italy, May 18, 2015, pp. 53-54. [Scholar Google]
- Cherkaoui, 0.; Elmoznine, R.; Boussu, F.; Cochrane, C.; Abed, A. The creation and implementation of a piezo-resistant sensor utilizing PEDOT: PSS applied to natural fiber from Sisal for 3D warp interlock fabric monitoring. Mater. Sci. Eng., IOP Conf. Ser. 2020, 827, 012019. [Scholar Google] [Cross Reference]
- 15. Single-Layer Pressure Textile Sensors with Woven Conductive Yarn Circuit: Kim, G.; Vu, C.C.; Kim, J. 2020; Appl. Sci. 10, 2877. [Scholar Google] [Cross Reference]
- Id. Park, S. and Jayaraman, S. Wearable Electronic Systems for Smart Textiles. MRS Bull. 2003, 28, 585-591. [Scholar Google] [Cross Reference]
- 17. Improving Wearable Technology to Improve Quality of Life, Park, S., Jayaraman, S. 2003; IEEE Eng. Med. Biol. Mag. 22, 41W8. [Scholar Google] [Cross Reference]
- Koncar, V. Smart Textiles for Applications in Monitoring and Measurement. Elsevier, Amsterdam, The Netherlands, 2019; Smart Textiles for In Situ Monitoring of Composites, pp. 1-151. 978-0-08- 102308-2 is the ISBN. [Scholar Google]
- Wearable electronics and smart textiles: a critical review by Stoppa, M. and Chiolerio, A. 2014; 14, 11957-11992. Sensors. [Scholar Google] [Cross Reference] [Version Green]
- 20. Rambusek, L. Textronics: Definition, Design, and Properties of Fibrous Organic Field Effect Transistors. Doctoral Thesis, University of Gent, Gent, Belgium, 2014. [Scholar Google]